



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/634,653

08/05/2003

Randall Lashinski

MITRAL.ICP3C2

6365

30452

7590

11/30/2006

EDWARDS LIFESCIENCES CORPORATION
LEGAL DEPARTMENT
ONE EDWARDS WAY
IRVINE, CA 92614

EXAMINER

ISABELLA, DAVID J

ART UNIT

PAPER NUMBER

3738

DATE MAILED: 11/30/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/634,653

Applicant(s)

LASHINSKI ET AL.

Examiner

DAVID J. ISABELLA

Art Unit

3738

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 27 June 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-3,6 and 12-14 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-3,6 and 12-14 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: _____

Response to Amendment

1. The amendment filed June 27, 2006 has been entered. Claim 1 has been amended to include the limitation of “the implant including a guidewire lumen adapted to slidably engage a guidewire”. Claims 1-3,6 were previously presented. Claims 4,5,7-11 have been cancelled and claims 12-14 have been newly added.

Response to Arguments

2. Applicant’s arguments, see pages 5 and 6 of the amendment filed June 27, 2006, with respect to the rejection(s) of claim(s) 1-3 and 7-11 under 35 U.S.C. § 103(a) have been fully considered and remain unpersuasive.

The newly added limitation of “the implant including a guidewire lumen adapted to slidably engage a guidewire” as broadly worded fails to distinguish over the embodiment as shown in figure 4e of Vidlund et al. Though the lumen 108 of Vidlund et al is not intended for engagement with a guidewire, structurally, the lumen is identical to that as broadly claimed and is capable to be slidably engage with a guidewire.

With respect to applicant's arguments directed to “flexibility” of the implant applicant's attention is directed to paragraphs [0117] and [0122] as well as figures 4h and 4i.

Claim Rejections - 35 USC § 103

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

4. Claims 1-3,6 are rejected under 35 U.S.C. 103(a) as being unpatentable over Vidlund et al. (USPAP 2003/0130731, as cited in applicant's IDS) in view of Liddicoat et al. (USPN 6,790,231, as cited in applicant's IDS) and Alferness et al. (USPAP 2003/0105520, as cited in applicant's IDS).

Vidlund et al. discloses a system for remodeling a mitral valve annulus with all the elements of claim 1, but is silent to the implant being reversibly movable between the first and second configurations and a control on the catheter for reversibly transforming the implant between the first and second configurations. See paragraph [0125] for a delivery catheter, and paragraph [0124] for an implant (110h) that is movable between a first, flexible configuration (Figure 4h) for delivery to a site adjacent the annulus of the mitral valve, and a second configuration (Figure 4i) for remodeling the mitral valve annulus. When the wire actuation mechanism (90) is pulled proximally, the distal end of the implant (110h) retracts, the implant changes shape to the second remodeling configuration, and the implant becomes rigid due to the tension created in the wire. Because the implant (110h) is implanted into the coronary sinus using catheter-based delivery techniques, it is clearly implied that the implant is detachably carried by the delivery catheter in some way. Liddicoat et al. teaches an apparatus for reducing mitral regurgitation wherein a wire (54) is pushed and pulled to reversibly

Art Unit: 3738

move an implant body (50) to a configuration that forces the posterior annulus anteriorly from within the coronary sinus, which improves leaflet coaptation and reduces mitral regurgitation. See Figures 3-5 and column 5, lines 37-53. It would have been obvious to one of ordinary skill in the art at the time of applicant's invention to look to the teachings of Liddicoat et al. to modify the system of Vidlund et al. by making the implant (110h) reversibly movable between the first and second configurations by pulling and pushing the wire actuation mechanism (90) in order for the implant (110h) to force the posterior annulus anteriorly from within the coronary sinus, and thereby improve leaflet coaptation and reduce mitral regurgitation. Alferness et al. teaches a system for effecting the mitral valve annulus geometry wherein an implant (30) is detachably carried by a delivery catheter (52) having a lumen (54) by being slidably received in the lumen (54). The implant (30) includes first anchor (32), second anchor (36) and a cable (34). The cable (34) has a coupling (38) that is coupled to the coupling (40) of a tension cable (42) disposed within the delivery catheter (52). When the tension cable (42) is pulled proximally, tension is applied to the cable (34) and the geometry of the mitral valve annulus is effected. The catheter (52) and tension cable (42) with coupling (40) are capable of being removed to complete the deployment process. See Figures 2-5 and paragraphs [0034], [0038], [0040] and [0041]. It would have been obvious to one of ordinary skill in the art at the time of applicant's invention to look to the teachings of Alferness et al. to modify the system of Vidlund et al. such that the implant is deployed into the coronary sinus using a similar technique. By including a coupling to the actuation mechanism (90), a control on the catheter in the form of a tension cable with

coupling can be used to reversibly transform the implant between the first flexible configuration and the second remodeling configuration by creating or releasing tension in the actuation mechanism (90) by pulling or pushing. After the desired configuration of the implant (110h) is achieved, the delivery catheter (52) and control can be removed.

Claim 2, see Figure 4i for the implant comprising an arc when in the remodeling configuration.

With respect to claim 3, Vidlund et al. does not expressly disclose that a best-fit constant radius curve corresponding to the arc has a radius within the range of from about 10mm to 20mm. However, according to Figure 4i and paragraph [0125] the arc of the implant (110h) of Vidlund et al. in the second remodeling configuration will have a radius sized to remodel the mitral valve annulus. Because the arc of the implant of applicant has a radius between 10-20mm in order to remodel the mitral valve annulus, it is obvious that the arc radius of Vidlund et al. will also fall within the required range of claim 3.

5. Claim 6 is rejected under 35 U.S.C. 103(a) as being unpatentable over Vidlund et al., Liddicoat et al. and Alferness et al. as applied to claim 1 above, and further in view of Solem et al. (USPN 6,210,432, as cited in applicant's IDS).

Vidlund et al., as modified by Liddicoat et al. and Alferness et al., discloses a system for remodeling a mitral valve annulus with all the elements of claim 1, but is silent to the device further comprising a coating on the implant, as required by claim 6. Solem et al. teaches a device for the treatment of mitral insufficiency, wherein the

device is coated with heparin in order to avoid thrombosis in the coronary sinus, thus reducing the need for aspirin, ticlopedine or anticoagulant therapy. See column 5, lines 14-17. It would have been obvious to one of ordinary skill in the art at the time of applicant's invention to look to the teachings of Solem et al. to modify the implant of Vidlund et al. by including a coating of heparin on the implant (110h) in order to avoid thrombosis in the coronary sinus, thus reducing the need for aspirin, ticlopedine or anticoagulant therapy.

6. Claims 12-14 are rejected under 35 U.S.C. 103(a) as being unpatentable over Vidlund et al., Liddicoat et al. and Alferness et al. as applied to claim 1 above, and further in view of Langberg et al. (USPAP 2002/0103533, as cited in applicant's IDS).

Vidlund et al., as modified by Liddicoat et al. and Alferness et al., discloses a system for remodeling a mitral valve annulus with all the elements of claim 2, but is silent to the implant comprising a compound curve when in the remodeling configuration, as required by claim 4. See paragraph [0124] for the elongated body (110h) having a final shape with an increased radius of curvature in some regions and a decreased radius of curvature in other regions. Langberg et al teaches using a thumbwheel member for actuating a rotational coupler for applying tension to the flexible member. (see figures 10-13. It would have been obvious to one of ordinary skill in the art at the time of applicant's invention to look to the teachings of Langberg et al, to modify the device of Vidlund et al. to provide for selective actuation and tensioning of the flexible wire to effect the final shape of the such that the shape will apply a force to a

discrete portion of the atrial wall of the coronary sinus to reshape the mitral valve annulus in treating dilated cardiomyopathy.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).


A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to DAVID J. ISABELLA whose telephone number is 571-272-4749. The examiner can normally be reached on MONDAY-FRIDAY.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, CORRINE MCDERMOTT can be reached on 571-272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3738

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



DAVID J. SABELLA
Primary Examiner
Art Unit 3738

DJI
11/24/2006